Contents

Introduction 4

Section 1: Traction background and principles 5

Historical background 5

Principles of traction 5

Care of traction apparatus 6

Knots 7

Anatomical terms 9

Section 2: Core care plans for nursing adults and children/young people 10

Adult template for care of patients undergoing skin traction 10

Child/young person model for care 12

Section 3: Traction types, applications and care plans 14

Skin traction (formerly Pugh's Traction) 14

Thomas splint traction 18

Hamilton-Russell traction 26

Skeletal Hamilton-Russell traction 28

Cervical traction 30

Halo traction 30

Gallows traction 36

Slings and springs suspension 39

Kendrick Traction Device 41

Traction at home: Thomas splint 43

Glossary of terms 46

References 48

Further reading 49

Figure 1: Stage 1

Pass the end of the cord twice around the item to be secured, leaving plenty of cord to enable you to complete the knot. You can always trim the excess cord later.

Figure 2: Stage 2

Pass the end of the cord around the length

BACK TO CONTENTS

Child/young person model for care

Trust logo here

Before you complete this care plan the child/young person and parents/carers must be asked for their agreement (consent) to care and treatment.

Date and time commenced:	Print name:	Signature:

Nursing care action	Rationale
Explain the procedure to the child/young person and family.	If traction is to be effective it is essential that tho the cat

Nursing care action	Rationale
If able and appropriate (ie, not for a child/ young person in Gallows) Encourage the child/young person to lift themselves up using a monkey pole to relieve pressure on the sacrum. Apply padding to bony areas if required.	As above.
Encourage the child/young person to perform plantar/dorsi-flexion exercises to both feet to encourage venous return (dependent on age/ weight of child/young person). (Not required for a child/ young person in Gallows traction.) Older children/young people should be	

To ensure effective application and care for all types of traction covered in Section 3, please refer first to:

- Section 1 on Care of the traction apparatus and Knots
- Section 2 , Adult core care plan and Child/young person core care plan.

Skin traction (formally Pugh's Traction)

The principles of skin traction are the basis for both Gallows traction and Thomas splint traction. When using basic skin traction, traction tapes are applied to the lateral and medial aspects of the skin on the lower limb, and with the attached weight hung freely over the bed a generating pulling force is achieved. It is used to create a traction force over a large area of skin which is then transmitted via the soft tissues to the bone.

Skin traction is often described as either 'adhesive' or 'non-adhesive'. Adhesive skin traction is used infrequently and only in exceptional circumstances.

This guidance can be used to apply skin traction to all age groups. Variation in the way this modality is applied may exist in clinical practice.

The application of skin traction should be carried out by at least two health care practitioners (HCPs) who are trained in the procedure.

To safely apply skin traction.

Equipment:

- one adhesive or non-adhesive traction kit
- crêpe bandages (if not in the kit)
- padding (if not integral to the kit)
- tape
- scissors.

If you are using counter traction, you will need:

- traction cord
- weights and carriers
- balkan beam
- cross bars
- two pulleys
- s ars

Action	Rationale
Suspending the limb:	
different methods exist to suspend the limb in the extensions. Choice depends on the type of bed and the traction equipment available in the hospital	Facilitates the patient's mobility in the bed. Reduces the potential for the development of pressure sores.
weight may be applied using a separate length of traction cord as per patient's notes	To ensure limb is maintained in alignment.
elevate the foot of the bed.	To provide counter traction.

Practitioners should:

- remove and check bandages every 24 hours and whenever necessary
- not remove the extensions unless there are clinical indications
- examine the skin integrity of the patient, and report any signs of skin damage to a nurse or doctor
- check the extensions for wrinkles and adhesion and change only if necessary
- check the patient's limb for any neurovascular change or calf pain, and report this to the nurse or doctor.

If using counter traction, practitioners should:

- always ensure the weights are connected securely and the traction cord and weights are checked at least daily, but usually as often as the patient is checked
- assess that the weight on the traction given is suitable. If the patient is moving either up or down the bed constantly, then either the weight is wrong or the amount of elevation. Inform the senior nurse and doctor about this and record the information in the patient's notes
-

Thomas splint traction

The Thomas splint is used in conjunction with skin traction or skeletal traction to immobilise and position fractures of the femur. It can be used in all age groups and involves application of a long leg splint with a hoop or ring that extends beyond the foot which can then be fixed, for example, to the end of the bed or suspended in a balanced system using skeletal or skin traction (Clarke and Santy Tomlinson, 2014). Application of a Thomas splint can be upsetting for the patient but most often reduces pain.

These guidelines can be used to apply Thomas splint traction to all age groups. Variation in the way this modality is applied may exist in clinical practice.

The application of Thomas splint traction should be carried out by at least two health care practitioners who are trained in the procedure.

To safely apply Thomas splint traction.

Equipment:

- Thomas splint: traditional splint with complete ring or adjustable splint with correct ring size and length (Figure 6).
- velcro (Figure 5) or calico slings
- adhesive skin extension set
- two crêpe bandages
- tapnt: <</L7.839 en-GBg(w)13.8 9567 94 23a9.5 (j)s(t)20.7ujen1 81 83.6 (-10.7 p(a)9. (en-GB)/MCID 795 >>

Action	Rationale

Action	Rationale
Apply skin extension, making sure that tape is placed to medial and lateral aspects of the limb. Ensure the skin extension tapes are free from creases and the foam padding is over the malleoli.	Protects bony prominences against incidence of pressure sores. Prevention of potential blister formation.
Leave sufficient room between the patient's foot and the end of the skin extension.	Allows for plantar flexion of the foot, maintaining normal ankle range of movement.
Bandage the skin extension to the patient's lower leg, using crêpe bandages. Avoid tight bandaging over the fibula head. Avoid bandaging over the malleoli and the Achilles' tendon (Draper and Scott, 1996). In children/young people, the bandages may go to the top of the leg. Secure the ends with tape. Leave knee area exposed. The position of the extensions both laterally and medially can be used to control rotation of the limb, ie, placing the lateral extension superiorly and the medial extension inferiorly, external rotation will be corrected.	

Action	Rationale
Place the leg on the splint, while still maintaining manual traction and limb support. A small piece of gamgee is placed behind the knee. The ring should fit snugly into the groin and up against the ischial tuberosity.	To maintain the traction pull. To allow slight flexion of the knee. To allow for fixed traction to be applied.
Secure the traction cords from the skin extension to the end of the splint, passing the outer cord over the lateral bar of the splint and the inner cord under the medial bar. Apply a windlass (the use of two spatulas is not best practice).	To allow application of fixed traction to the limb. To prevent external rotation of the limb. Takes up the slack in the tied ends and exerts an increased traction force to the limb.

Action	Rationale
Reassess the patient's neurovascular status. If a child/young person, crêpe bandages can be applied over the top of the splint. Place gamgee over thigh to help further reduce fracture along with soft tissues and muscles.	
If a child/young person, crêpe bandages can be applied over the top of the splint.	Prevents the child/young person from disturbing the underlying extension set and

Action	Rationale
Ensure the patient has an AP and lateral X-ray of the limb.	To check the position of the fracture, confirming its reduction. Allow for further adjustments to be made if required.
Patient's fracture is immobilised and patient can be safely transferred to the ward. Complete neurovascular observations on the affected limb and document.	 Adult and child/young person's care on page 10 and page 12.
Suspending the splint (see Figures 7 and 8): Different methods exist to suspend the limb in the splint, depending on the type of bed and the traction equipment available in the hospital. Weight may be applied using a separate length of traction cord. Elevate the foot of the bed.	Facilitates the patient's mobility in the bed. Reduces the potential for the development of pressure sores. To reduce pressure in the groin from the ring. To provide counter traction.

Figure 7: Suspended Thomas splint

Figure 8: Suspended Thomas splint

Hamilton-Russell traction

Hamilton Russell is a balanced traction system using vectors to effect a pull along the long axis of the femur. It is used to maintain the joint space at the hip, manage fractures of the acetabulum and support fractures of the shaft of the femur. Traction can be applied using below-knee skin traction or a skeletal pin (Clarke and Santy Tomlinson, 2014).

This guidance can be used to apply Hamilton-Russell traction to all age groups. Variation in the way this modality is applied may exist in clinical practice, however.

The application of Hamilton-Russell traction should be carried out by at least two health care practitioners who are trained in the procedure.

To safely apply Hamilton-Russell traction via the skin.

Equipment:

- adhesive/non-adhesive traction kit
- two crêpe bandages (if not included in the kit)
- cotton wool padding with non-woven cover
- tape
- traction cord
- weights and carrier
- 15cm wide sling
- pillow(s)
- Balkan beam with a cross bar above the knee, and two extension bars with crossbars at the foot end of the bed
- three pulleys
- scissors
- spreader plate with pulley attached.

Figure 9: The application of Hamilton-Russell traction using one sling

Skeletal Hamilton-Russell traction

To safely apply Hamilton-Russell traction via a skeletal pin through the upper tibia.

Equipment:

- skeletal pin
- pin covers
- stirrup
- U loop with pulley attached
- bandage
- large bulldog clips or safety pins
- traction cord
- weights
- pillow(s)
- Balkan beam with a cross bar above the knee, and two extension bars with crossbars at the foot end of the bed
- three pulleys.

Action	Rationale
Explain the fitting of the traction to the patient and family/carers.	To gain informed consent and enable the patient to have full understanding of the need for the tra

Practitioners should:

- maintain the lower leg parallel to the bed frame
- maintain the degree of abduction prescribed by the medical staff. This is controlled by the position of the pulleys on the crossbars at the foot of the bed
- maintain the amount of flexion at the hip joint prescribed by the medical staff. This may have to be compromised by other clinical considerations. A fractured shaft of femur normally requires 20 degree flexion (Larson and Gould, 1978)
- the patient may be tilted laterally, but should still be aligned to the mid-line and the hip abduction and rotation controlled by the mechanics of the pulley system
- document procedure and care in the patient's notes.

Cervical traction

"Cervical traction is a non-invasive method for the reduction and stabilisation of spinal fractures and dislocations as the method of choice for patients unsuitable for surgery or for whom surgery is not a priority" (Harrison and Ash, 2011 p.565).

Halo traction

Halo traction is the application of skeletal traction to the head by means of a halo device.

A 'halo vest' provides traction and support which allows for proper healing of the damaged or diseased spinal region. A halo vest includes a 'halo ring' that is secured to the skull with skull pins and a rigid, fleece-lined vest. The vest is attached to the halo ring with adjustable metal rods or struts. Together the apparatus provides stability to the cervical column while allowing the patient to be mobile.

Halo traction is applied to the patient in theatre or the emergency department by a trained team and radiographers.

Some halo rings and vests are MRI compatible. Check with the MRI department if an MRI test is required.

Halo traction is indicated in adults for definitive treatment of cervical spine trauma including:

- occipital condyle fractures
- occiptio-cervical dislocation
- stable typ(H)5.6 4pI Aiobioe326 >>BDC B13.30.7 (()210.2 144.4099 Tm [(s)12.2 (t)6.9 (a)7.8 (b)JTm or d

Halo traction is indicated in children/young people for:

- atlanto-occipital disassociation
- Jefferson fractures (burst fracture of C1)
- atlas fractures
- unstable odontoid fractures
- persistent atlanto-axial rotatory subluxation
- C1-C2 disassociations
- sub axial cervical spine trauma
- preoperative reduction in patients with spinal deformity.

Contra-indications:

- cranial fractures
- infection
- severe soft-tissue injury
- especially near proposed pin sites
- polytrauma
- severe chest trauma
- barrel-shaped chest
- obesity
- advanced age
- poor mental capacity.

To apply halo vest to immobilise the spine by restricting head, neck and upper body movement to promote healing of the damaged or diseased spinal region.

When spinal cord lesions prevent messages of painful stimuli (most commonly, distended bladder or constipation) from reaching the brain, an autonomic nervous system reaction is initiated in the body. This is most common in spinal injuries above T6 and presents as episodes of high blood pressure, throbbing headaches, profuse sweating, flushing of the skin (above the level of the spinal lesion), and nasal stuffiness.

This is a medical emergency, so recognising and treating the earliest signs and symptoms efficiently is crucial.

Action	Rationale
Vest care: Halo vest, straps/buckles and bolt integrity will be assessed once each shift and PRN. Loose bolts or worn straps should be reported to medical staff. Buckles should not be released except for skin care or vest maintenance. The sheepskin should be kept dry at all times. If soiled or wet, notify the practitioner who is trained in halo vest application, to have the sheepskin changed. If edges are only slightly damp, expose to air and reassess. Do not add any padding under the vest for comfort. If the vest does not fit properly, notify the practitioner for assessment and adjustment as required Do not use any supports under the head. Assess patient's swallowing. If a patient is having difficulties swallowing, a swallowing evaluation should be completed by a speech-language specialist. ea (e)13.7 (t)2t 7et25r7 (ng)-6.8 (. is)-2.2 (t)-19d	

Action	Rationale
Skin and vest care: Patients wearing a halo vest are at increased risk for skin breakdown and should be assessed for pressure areas once every shift at a minimum, with special attention to bony prominences. Patients with neurological deficits may not be aware of pressure areas and require increased visual inspections. If patient is non-compliant, at least two staff must be	

Action	Rationale
Repositioning and mobilising:	

Gallows traction

Gallows traction is used in very young children (usually <18 months) to treat a fractured shaft of femur, or may be used to stretch the soft tissues pre-operatively in the treatment of developmental dysplasia of the hip (DDH). Gallows traction is appropriate for a non ambulant child and weighing no more than 12-16 kgs. Compartment syndrome GgG.8 (7.2 (f)-4.1)2 (mur)15.7 2y < (p)-1.8 (p)-1.87 2y < (p)-1.5 (d w)16.c (n -8.4 e t)15.5 (or6 a)]2 (he h)7.5

Action	Rationale
Re-apply the bandages daily. Check the area of skin that is visible for soreness.	To prevent bandages wrinkling and causing potential tightness, which would cause skin problems and potentially vascular compromise. To examine the skin without disturbing the adhesive traction.
Continue to monitor the circulation two-hourly, for the first 48 hours.	Check for neurovascular deficit.
Check that pressure is not applied over the fibula head and that the child is moving their toes and feet.	To avoid pressure on the common peronealnerve (also known as the lateral popliteal nerve), to prevent the development of foot drop.
Provide age-appropriate toys to entertain the child in conjunction with the play staff.	To encourage the child's mental, physical and social development.
Encourage the parents to participate in their child's care eg, feeding, bathing, nappy changing, traction and bandage checks.	To involve them in the care of their child.
When the skin extensions need to be removed, they should be soaked off in a bath. The use of plaster- dissolving solution is also useful.	To reduce the risk of skin irritation.

See: uhs.nhs.uk/Media/UHS-website-2019/Patientinformation/Childhealth/Gallows-traction-patient-information.pdf

Slings and springs suspension

‘Slings and springs’ is not true traction, as the aim is to hold the leg in suspension with no true counter traction. There is no empirical evidence to support its use.

However, Maqueen et al., (2010) report slings and springs traction as a good method for restoring range of motion gradually to an irritable hip, by abducting the effected leg, and secondly to rest the hip joint in a flexed position whilst allowing active movement in a child presenting with Perthes disease. Slings and springs is also used to allow joint mobilisation while a patient is non-weight bearing following femoral and tibial osteotomy.

Equipment:

-

Action	Rationale
Give a full explanation to the parents and child/young person. Use visual aids such as pictures and DVD. Allow time to answer questions. Gain verbal consent for the procedure.	To ensure the parents are fully informed, that they understand the reason for the traction and consent to the procedure.
Overhead beam to be in position above either right or left hip (a degree of abduction may be ordered by orthopaedic consultant). With child/young person positioned on bed, attach a pulley above the child's affected thigh and calf.	To attach slings and springs, provide comfort and encourage range of movement.
Thread traction cord singular through each pulley which is attached to spring and then sling Spring and sling then drops vertically from overhead beam and may need to be adjusted to suit each child/ young person.	To suspend the leg at a comfortable height.
Ensure knots are securely tied and check that the cord does not fray.	Sa46.0en-GBtf y]TJ EMC /Span <</Lang (en-GB)/MCID 1753 4>BDC

sanao (.)]TJ EM
trte rtong ildng

BACK TO CONTENTS

Figure 12: Slings and springs suspension

Practitioners should:

- t

Figure 13: Kendrick traction device equipment



Figure 14: Kendrick Traction Device applied to mannequin

Figure 15: Ankle hitch



Traction at home: Thomas splint

Conservative management of femoral shaft fractures with Thomas splint traction in children/young people is a treatment option, but the treatment time is six to eight weeks. This length of hospital stay makes this type of treatment as an inpatient unacceptable

1. Slings, padding and all traction bandages checked/ changed as required before final X-ray. Windlass secured with tape (see Figure 17 below).
2. X-ray carried out and reviewed to ensure early callus formation is seen.
3. Weights removed. Bed put in flat position. Plaster cast applied over bandages (thin layer only so it is easily removed in the home with plaster shears) to stop child removing bandages and padding.
4. Child stays in hospital for a further 24 hours to ensure that ring does not cause groin pressure following weight removal. If skin pressure does occur, turns on windlass need to be reduced and if there is a significant change in traction then a further X-ray check may be required. If child reports increased pain in this period it indicates fracture not stable, so child is not ready for transfer home.

Figure 17: Securing the windlass

Parents and carers need this information to plan care at home. The times are estimates and every child will have an individual response to treatment.

- The total fracture healing time in weeks is usually the age of the child plus two.
- Early callus formation one third of total healing time.
- Rehabilitation time two to four weeks (that is, time before child/young person is walking independently following removal of traction).
-

Care is now provided by parents/carers. They should be provided with a telephone contact number which they can ring for advice at all times. At home, the child/young person should be kept on bed rest.

The child/young person may sit up for meals and for a short period afterwards. A child/young person who is allowed to sit up for too long will experience skin pressure problems, therefore for the majority of the time, they should not be made comfortable sitting up – most children/young people will sit themselves up, but if they aren't supported will lie down again. Lying pro0.6 (w)8.5 EMC (r)9.m [(B<C T* [(i)9.822i175) 7ng.1 (rs)4(h)-2. (r)8.9 (i)5.1.1 (tn)T

Thild/yor as0.6 (l)922i171.D 1910 7 (o -4.7 0)-4.7 (/)k1.D 1 (i)19.7e tdefiniA (e)2.1 (r)8.9 (is)12.5.2 (he)2 .5 (i)

Abduction – movement of a limb from the midline of the body.

Adduction – movement of a limb closer to the midline of the body.

Anti-embolic – chemical or mechanical methods used to prevent blood clot formation.

Balkan beam – vertical poles at the head and foot of the bed joined by overhead horizontal poles. Both the upright and horizontal poles can have adjustable cross bars positioned as required for the attachment of pulleys.

Bohler stirrup – a 'U'-shaped stirrup with special ends which fits onto a Steinmann or Denham pin with a loop at one end for attachment of traction cord.

Capillary refill – assessment of circulation return to small blood vessels.

Cast padding – used to protect bony prominences from pressure, to reduce risk of pressure ulcers from developing.

Common peroneal nerve (lateral popliteal nerve) – nerve which winds around the head of the fibular near the knee joint.

Compartment syndrome – a condition caused by increased pressure within a compartment of a limb which leads to reduced oxygenation of the tissues and eventually death of those tissues. Plantar flexion of the foot with inversion will increase pressure within the calf alone even without the application of a cast or traction.

Counter traction – a force acting in the opposite direction to the pull or line of traction on the limb or body.

Deep vein thrombosis (DVT) – a blood clot that forms in the veins called a thrombus. Clot can then break off to form an embolus eg, pulmonary embolus, which can be fatal.

Denham pin – a rigid stainless steel pin with a short threaded area; usually 4-6mm in diameter and of varying lengths.

Distal – part furthest away from the centre of the body.

Dorsalis pedis – pulse pal973 TD [(D)-5 p9iiilistVle ooon tcan t – pulVu – a forculsl9.12(e)7.3 (i)9.1 <</La.4

Hyperextension – excessive extension of a joint beyond normal limits. Can be a normal finding in some patients' knees.

Nissen stirrup – similar to the Bohler stirrup but has multiple positions to attach traction cord to give some control over rotation of the leg.

Non-union – no healing or very slow healing of a fracture.

Optimum – the best possible in any given set of circumstances.

Plantar flexion – relates to the sole of the foot, the plantar side, when the toes point downwards. The opposite of dorsiflexion.

Proximal – part nearer to the centre of the body.

Radial pulse – pulse palpable at the wrist joint just below the base of the thumb.

Recumbent – lying down flat on the back or sides.

Semi-recumbent – as recumbent, but with the trunk slightly raised.

Skin extensions – either adhesive or non-adhesive, available in pre-made packs; applied to inner and outer side of limbs to allow traction to be applied.

Slings for Thomas splint – non-stretchy material placed across the bars of the splint and anchored on the outer side within safety pins.

Spreaders – a device used to allow traction cord to slide without compression on the patient's limb.

Steinmann pin – a rigid stainless steel pin, 4-6mm in diameter of varying lengths.

Thomas splint – two longitudinal bars, terminating in a w-shaped extension bow at the distal end. Proximally, a leather ring is set obliquely at an angle of 120 degrees to the metal bar. Rings can be full, split or half.

Thrombus – the formation, presence or development of a clot or thrombus in a vein.

Traction cord – strong cord which can be tied securely without slipping

Rockwood CA and Green EP (2010) Fractures in Adults, Volume two (Seventh edition), New York: Walters Kluwer/ Williams Lipponcot, Wilkins.

Royal College of Nursing (2002) A Traction Manual. London: RCN.

Royal College of Nursing (2011) Guidance on pin site care: Report and recommendations from the 2010 consensus project on pin site care , London: RCN. Available at: www.rcn.org.uk/professional-development/publications/pub-004137 (accessed 17 June 2021).

Royal College of Nursing (2014) Peripheral neurovascRon: RCN. Available at: up9ofevelopmeenpubca s d0.7

Royal C ege
f Nurs n